

APAVE SUDEUROPE SAS
LABORATOIRE DE LYON TASSIN
Accredited body by COFRAC n° 1-1461
Essais Sur Produits
177, route de Sain-Bel
69160 TASSIN LA DEMI-LUNE

BEEWAIR
3 Allée de la Teppe
BP60031
71012 MACON France

SYNTHESIS DOCUMENT

PRODUCTS CONFORMITY ASSESSMENT

1 SCOPE OF THIS DOCUMENT

The present document constitutes the synthesis of the APAVE report n°8185185 which gives the details of inspections and tests achieved on the product.

2 CONCERNED PRODUCT SPECIFIED FEATURE

- | | | | |
|------------------|--------------------------------|----------------------|-----------------------------|
| • Designation : | AIR treatment system DBD-Lyse. | | |
| • Manufacturer : | BEEWAIR | • Voltage : | AC 120V 1.25A T 60Hz
40W |
| • Model : | BW60L | • Supply connection: | Appliance inlet |
| • Reference | 109 086 and
109 085 | • Serial number : | 47C / 001 006 |

3 SCOPE OF THE MISSION (to see details in report)

Conformity Assessment of the product and its technical file to the following test specification :

- ⇒ US National standard ANSI/AAMI ES60601-1: 2005 + A2:2010 +A1: 2012
- ⇒ Canadian National standard: CAN/CSA-C22.2 No. 60601-1:14


CONCLUSION OF THE REPORT

The conclusion of the report shows that the examined product is in compliance with the prescriptions of the reference documents.

Tassin the 16th December 2014

The test engineer

Limit of validity of this document : 2 years



Sylvain Gerin

This document is not a report established under the title of a Notified body, it is a synthesis of the mission achieved by APAVE and only the complete French report maybe used as reference. It doesn't substitute itself to the DECLARATION OF CONFORMITY issued by the manufacturer, according the market regulation of the EUROPEAN UNION and cannot be use as a guarantee of the production conformity to the type
Its reproduction is only allowed as a complete photographic facsimile